

Declaration of CE Conformity

Manufacturer: Biochrom Ltd
Address: 22 The Science Park, Cambs, CB4 0FJ, UK
Type of Equipment: Laboratory Equipment
Product description: **Microplate Reader Anthos Zenyth 340R, 340RT, 340S, 340ST, 200RT, 200ST**

The described products meet the essential requirements set out in the following Directives

73/23/EEC & 89/336/EEC

Standards against which conformity is assessed:

EN61010-1: (1993)

Safety requirements for electrical equipment for measurement, control and laboratory use
EN61326-1 (1997 +A1:1998) Electrical equipment for measurement, control and laboratory use. EMC requirements

EN55011 (1998 + A1:199) Limits and methods of measurement of radio disturbance characteristics of industrial, scientific and medical (ISM) radio-frequency equipment

IEC61000-4-2: (a995 + A1:1998) Testing and measurement techniques - Electrostatic discharge immunity tests.

IEC61000-4-4: (1995) Testing and measurement techniques - Electrical fast transient/burst immunity test.

IEC61000-3-2: (2000) Limits for harmonic current emissions

IEC61000-6: (2004) Generic standards – Immunity for residential, commercial and light-industrial environments

EN55022 (1994 + A:2000 + A2:1997) Information technology equipment. Radio disturbance characteristics. Limits and methods of measurement

Original Issuer: Asys Hitech GmbH

Place, Date: Eugendorf, November 5th, 2007

A handwritten signature in black ink, appearing to be "S P Marriott", written over a faint horizontal line.

Signature:

S P Marriott

R&D Director, Biochrom

This declaration confirms the agreement of the mentioned guidelines, comprises no assurance of features.
Please follow the safety notes of the included product information.